



inc. BIOLOGICAL CONSULTING SERVICES
OF NORTH FLORIDA, INC.

February 25, 2016

Icon Lifesaver Ltd.
Hall Chase, London Road
Marks Tey, Colchester
CO6 1EH, UK
+44(0)1206 580999

RE: Biological filtration efficacy test study of the provided Icon Lifesaver® cube filter units CUB5k_5-7; BCS IDs 1602122, 1602123, and 11602124.

To whom it may concern,

We have conducted the requested filtration efficacy study on the filter units received on February 12th, 2016. The experimental set up and challenge of the water filters was designed to evaluate the filters microbiological contaminant removal efficacy. The contaminant species and water parameters selected were based on client's request and guidance from NSF/ANSI P231 water purifier test protocol. The units' challenge parameters were selected to simulate operation of the filter units by personnel.

In the following pages, you will find a summary of the methodology used and the results of our analysis. Should you have any questions or concerns, please do not hesitate to contact me.

Best Regards,

George Lukasik, Ph.D.
Laboratory Director

- PAGE 1 OF 13 -

BCS LABORATORIES, INC. - GAINESVILLE
4609 NW 6TH STREET, STE. A, GAINESVILLE, FLORIDA 32609
TEL. (352) 377-9272, FAX. (352) 377-5630

WWW.MICROBIOSERVICES.COM

FL DOH #E82924, ISO/IEC 17025:2005 L2422 (L-A-B), EPA# FL01147

FILE: ICON LIFESAVER CUBE FILTER TESTING BCS 1602122-124 02.17.2016

THIS REPORT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN CONSENT OF BCS LABORATORIES



Test Article(s):

On February 12th, 2016, 4 cube filter units were received from Icon Lifesaver. The four cube filter units had the designation CUB5k_5-8, the units were issued BCS identifiers 1602122, 1602123, 1602124, and 1602125 respectively. BCS IDs 1602122, 1602123, and 1602124 were selected for the study and 1602125 was kept in reserve.

Study Date:

The study was initiated on February 16th, 2016 and completed on February 18th, 2016.

Performed by: David Sekora, M.S.

Analyzed by: David Sekora, M.S.

Study Supervisor: George Lukasik, Ph.D.

- PAGE 2 OF 13 -

BCS LABORATORIES, INC. - GAINESVILLE
4609 NW 6TH STREET, STE. A, GAINESVILLE, FLORIDA 32609
TEL. (352) 377-9272, FAX. (352) 377-5630

WWW.MICROBIOSERVICES.COM

FL DOH #E82924, ISO/IEC 17025:2005 L2422 (L-A-B), EPA# FL01147

FILE: ICON LIFESAVER CUBE FILTER TESTING BCS 1602122-124 02.17.2016

THIS REPORT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN CONSENT OF BCS LABORATORIES



Physical parameter measuring devices and critical equipment utilized:

Equipment and Measurement Parameter	Manufacturer	BCS Lab ID
Balance	Sartorius Laboratory Instruments	BL-4
Epi-fluorescence microscope	Olympus BH-2	MIC-3
Digital Colorimeter; DPD-02	Hach DR 890	COL-03/DPD-02
Turbidity meter	Hach Turbidity Meter	TM-01
Alkalinity test kit	LaMotte	511220
Total hardness test kit	LaMotte	4911208
Incubator	Sanyo MIR-253	I-2
pH	SevenCompact pH/Ion pH meter, Model S220	PH-4
Conductivity/TDS	VWR Traceable Conductivity Meter 89094-958	CM-05
Timer	VWR Traceable 61161-346	T-10
Centrifuge	Eppendorf C-5702	C-12
Temperature	VWR NIST traceable IR Thermometer 33501-413	IR-5
4-Liter standardized graduated cylinder	Nalgene	GC-4L-A
Pressure regulator	Ingersoll Rand PR4021-200	(PR-1)
Pressure Vessel; 55 liter	Alloy Products 55L 67349008	PV-06
Pressure Transducer NIST 2 BAR	Sper Scientific PS100-5BAR	PM 29
NIST digital pressure meter 0-100 PSI	Omega DPG 1001B-100G	PM 27

BCS LABORATORIES, INC. - GAINESVILLE
 4609 NW 6TH STREET, STE. A, GAINESVILLE, FLORIDA 32609
 TEL. (352) 377-9272, FAX. (352) 377-5630

WWW.MICROBIOSERVICES.COM

FL DOH #E82924, ISO/IEC 17025:2005 L2422 (L-A-B), EPA# FL01147

FILE: ICON LIFESAVER CUBE FILTER TESTING BCS 1602122-124 02.17.2016

THIS REPORT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN CONSENT OF BCS LABORATORIES



Test Matrix; General Test Water 1:

General Test Water 1 (GTW1) was made up of the dechlorinated municipal water. Municipal water was dechlorinated by filtration through carbon block filters. Total dissolved solids, turbidity, and pH were measured and adjusted (if necessary) to NSF P231 guidelines. The pH of the water was 7.05, turbidity was 0.46 NTU, total dissolved solids were measured at 169 ppm, and Total Organic Carbon (TOC) was <1.0 ppm. Temperature was maintained between 22°C and 23°C. TOC analysis was conducted by XENCO Laboratories (Tampa, FL).

Test Matrix; Challenge Test Water 3:

Challenge Test Water 3 (CTW3) was prepared from dechlorinated municipal water and adjusted to NSF P231 guidance for total dissolved solids, turbidity, and pH. The pH of the test water was 8.96, turbidity was 31.4 NTU, total dissolved solids were measured at 1550 ppm, and Total Organic Carbon (TOC) was >10 ppm. Temperature was maintained between 4.1°C and 4.3°C. TOC analysis was conducted by XENCO Laboratories (Tampa, FL).

Challenge Species:

Bacteria: *Raoultella terrigena* ATCC ® 33257 reference stock culture was obtained from Microbiologics® (MN, USA) and maintained as per supplier's recommendations.

The lyophilized culture was hydrated and propagated on Tryptic Soy Agar (TSA,

- PAGE 4 OF 13 -

BCS LABORATORIES, INC. - GAINESVILLE
4609 NW 6TH STREET, STE. A, GAINESVILLE, FLORIDA 32609
TEL. (352) 377-9272, FAX. (352) 377-5630

WWW.MICROBIOSERVICES.COM

FL DOH #E82924, ISO/IEC 17025:2005 L2422 (L-A-B), EPA# FL01147

FILE: ICON LIFESAVER CUBE FILTER TESTING BCS 1602122-124 02.17.2016

THIS REPORT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN CONSENT OF BCS LABORATORIES



Neogen Inc., MI). Prior to the date of the study, a broth culture (Tryptic Soy Broth (TSB), Neogen Inc., MI) was started from a single colony. The culture was incubated at 36.5 ± 0.5 °C for 15-18 hrs. On the day of the study, the culture was centrifuged at 3K x G for 10 minutes and suspended in laboratory grade reagent water. Bacteria were enumerated following dilution in Phosphate Buffered Water (PBW, 3M, USA) by spread plating onto TSA. Standard Method 9215C (APHA, 2012) was used for the enumeration of *Raoultella terrigena*. Briefly, duplicate 0.1 and 1.0 mL samples of the filters' effluent and influent (10^{-3} dilution in PBW) were plated onto TSA. The plates were incubated at 36.5 ± 0.5 °C for 18-20 hours prior to colony enumeration.

Virus: Bacteriophage MS2 (ATCC 15597-B1; 30 nm RNA virus specific for *Escherichia coli* C3000 ATCC 15597) was used in this study as a surrogate for viral pathogens. The virus was cultivated to $>10^{10}$ plaque forming units (pfu)/mL in the laboratory prior to the challenge study. Bacteriophage stock was pre-filtered through a 0.22 μ m membrane filter (Millipore, USA). Titer was determined by performing 1,000 fold dilutions of bacteriophage stock in sterile PBW and enumeration as per laboratory standard methodology (SOP V-10). Bacteriophage stock was maintained at 4°C until the initiation of the challenge study. For the enumeration of MS-2 in collected samples, duplicate 0.1 and 1.0 mL aliquots of each of the collected samples were analyzed by an agar overlay plaque assay using *E. coli* ATCC 15597 as the host. Dilutions of high titer samples (filter influent) was performed in PBW prior to analysis. Plates were incubated

- PAGE 5 OF 13 -

BCS LABORATORIES, INC. - GAINESVILLE
4609 NW 6TH STREET, STE. A, GAINESVILLE, FLORIDA 32609
TEL. (352) 377-9272, FAX. (352) 377-5630

WWW.MICROBIOSERVICES.COM

FL DOH #E82924, ISO/IEC 17025:2005 L2422 (L-A-B), EPA# FLO1147

FILE: ICON LIFESAVER CUBE FILTER TESTING BCS 1602122-124 02.17.2016

THIS REPORT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN CONSENT OF BCS LABORATORIES



at 36.5±5°C for 18-20 hours prior to plaque enumeration.

Parasite surrogate: 3.0 micrometer Fluoro-Max Green Fluorescent Polymer Microspheres (Lot 43393) were obtained from Thermo Scientific (USA) and validated to the correct size using scanning electron microscopy (SEM, University of Florida, US). Three well slides (PTFE Slides, Electron Microscopy Sciences, USA) were used for sample mounting and enumeration under fluorescent UV microscopy (FITC Filter). Enumeration was conducted as per EPA1623.1 methodology. All collected samples were analyzed in duplicates at the minimum.

Challenge study Description / Methodology:

The provided filters were fitted with appropriate connections to the source of GTW1. Each unit was first conditioned with GTW1 as per manufacturer's instructions. Approximately 25 liters were passed through each filter prior to challenge. The starting pressure was adjusted to 4.5 PSI throughout the challenge study. The pressure was chosen as it was equivalent to pumping a cube filter unit 20-22 times with the valve closed. The flow rate varied between the units and was measured at 590-640 mL/min at 4.5 PSI. For initial challenge water preparation, aliquots of the challenge species were added to 40 liters of GTW1. The water was homogenized and an influent sample was removed and preserved for enumeration. Following the initial conditioning of the filters, each cube unit was emptied and filled with challenge water (GTW1 + challenge

- PAGE 6 OF 13 -

BCS LABORATORIES, INC. - GAINESVILLE
4609 NW 6TH STREET, STE. A, GAINESVILLE, FLORIDA 32609
TEL. (352) 377-9272, FAX. (352) 377-5630

WWW.MICROBIOSERVICES.COM

FL DOH #E82924, ISO/IEC 17025:2005 L2422 (L-A-B), EPA# FLO1147

FILE: ICON LIFESAVER CUBE FILTER TESTING BCS 1602122-124 02.17.2016

THIS REPORT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN CONSENT OF BCS LABORATORIES



species). Each of the units was sealed and pressurized. A minimum of 10 filter bed volumes were allowed to pass through each of the filters prior to collecting two consecutive 50 mL samples of the effluent. Pressure and time elapsed for the volumes collected were recorded using a validated measuring device. Upon completion of the GTW1 study, a second sample was taken from the remaining challenge water.

The filter challenge was then repeated using CTW3 in an identical manner. The flow rate varied between the units and was measured at 570-620 mL/min at 4.5 PSI. All collected filters' influent and effluent samples were assayed as per Standard Methods and Lab Standard Operating Procedures. All collected samples were analyzed, at a minimum, in duplicate for each sample volume and dilution. The respective percent reductions were determined based on the average concentration obtained from the duplicate samples of filters' influent and effluent samples at each specific challenge point.

- PAGE 7 OF 13 -

BCS LABORATORIES, INC. - GAINESVILLE
4609 NW 6TH STREET, STE. A, GAINESVILLE, FLORIDA 32609
TEL. (352) 377-9272, FAX. (352) 377-5630

WWW.MICROBIOSERVICES.COM

FL DOH #E82924, ISO/IEC 17025:2005 L2422 (L-A-B), EPA# FL01147

FILE: ICON LIFESAVER CUBE FILTER TESTING BCS 1602122-124 02.17.2016

THIS REPORT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN CONSENT OF BCS LABORATORIES



Study data are summarized in the provided table(s). The results presented pertain only to the study conducted on the test articles/samples/units provided by the client (or client representative). The study was authorized and commissioned by the client. The analytical results pertain only to the samples analyzed relating to the respective identifier number(s) indicated. The data provided is strictly representative of the study conducted using the material/samples/articles provided by the client (or client's representative) and it's (their) condition at the time of test. The study and data obtained under the laboratory conditions may not be representative or indicative of a real-life process and/or application. Positive, negative, and neutralization controls were performed as outlined in the method and as per Good Laboratory Practices. All analyses were performed in accordance with laboratory practices and procedures set forth by ISO 17025-2005 and NELAP/TNI accreditation standards unless otherwise noted. BCS makes no express or implied warranty regarding the ownership, merchantability, safety or fitness for a particular purpose of any such property or product.

- PAGE 8 OF 13 -

BCS LABORATORIES, INC. - GAINESVILLE
4609 NW 6TH STREET, STE. A, GAINESVILLE, FLORIDA 32609
TEL. (352) 377-9272, FAX. (352) 377-5630

WWW.MICROBIOSERVICES.COM

FL DOH #E82924, ISO/IEC 17025:2005 L2422 (L-A-B), EPA# FL01147

FILE: ICON LIFESAVER CUBE FILTER TESTING BCS 1602122-124 02.17.2016

THIS REPORT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN CONSENT OF BCS LABORATORIES



Project: Icon LIFESAVER® cube Efficacy Test
Sample(s): BCS 1602122, 1602123, and 1602124 received February 12th, 2016
Test: Filtration Efficacy – General Test Water (GTW) Type 1
Test Parameter: *Raoultella terrigena* (Bacteria), MS-2 Bacteriophage (virus), and 3.0 µM Fluorescent Microspheres as *Cryptosporidium parvum* oocyst surrogate
Test Date: February 16th, 2016

Challenge Species	Filter influent average concentration	Average concentration of the challenge species in the filters' effluent		
		CUB5k_5 BCS 1602122	CUB5k_6 BCS 1602123	CUB5k_7 BCS 1602124
Bacteria: <i>Raoultella terrigena</i> ¹	6.2 x 10 ⁵ cfu/mL	< 0.45 cfu/mL*	< 0.45 cfu/mL*	< 0.45 cfu/mL*
Virus: MS-2 Bacteriophage ²	5.4 x 10 ⁵ pfu/mL	5.7 pfu/mL	4.1 pfu/mL	7.2 pfu/mL
3.0 µM Fluorescent microspheres ³	4.1 x 10 ⁴ particle/mL	< 1.0 particle/mL*	< 1.0 particle/mL*	< 1.0 particle/mL*

¹ *Raoultella terrigena* (ATCC 33257) was obtained from ATCC and propagated on Tryptic Soy Agar (TSA, Becton Dickinson, USA). It is used to evaluate filters' bacterial removal efficacy. Bacteria was enumerated as colony forming units (cfu) following incubation at 36.5°C for 24 hours as per Standard method 9215C (APHA, 2012).

² Bacteriophage MS-2 (ATCC 15597-B1) was used as a model for human viruses. It is of similar shape and size to human enteroviruses and thus is used to determine filter's viral capture efficacy. It was enumerated using *E. coli* C3000 (ATCC 15597) as a host using the single layer plaque assay agar procedure as per EPA 1601.

³ Three micron green fluorescent latex microspheres (Fluoro-Max™ Green Fluorescent Microspheres 3.00µm, Thermo Scientific CA, USA) were used as surrogates for *Cryptosporidium* oocysts. It is used to determine filter's parasitic removal efficacy. The microspheres were enumerated by fixing onto 3-Well PTFE Slides (Electron Microscopy Sciences, USA) and viewing by UV fluorescence microscopy.

* No species were detected in the filter effluent for the total volume analyzed. Filter effluent samples were analyzed in duplicates at the minimum following collection.

BCS LABORATORIES, INC. - GAINESVILLE
 4609 NW 6TH STREET, STE. A, GAINESVILLE, FLORIDA 32609
 TEL. (352) 377-9272, FAX. (352) 377-5630

WWW.MICROBIOSERVICES.COM

FL DOH #E82924, ISO/IEC 17025:2005 L2422 (L-A-B), EPA# FLO1147

FILE: ICON LIFESAVER CUBE FILTER TESTING BCS 1602122-124 02.17.2016

THIS REPORT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN CONSENT OF BCS LABORATORIES



Project: Icon LIFESAVER® cube Efficacy Test
Sample(s): BCS 1602122, 1602123, and 1602124 received February 12th, 2016
Test: Filtration Efficacy – General Test Water (GTW) Type 1
Test Parameter: *Raoultella terrigena* (Bacteria), MS-2 Bacteriophage (virus), and 3.0 µM Fluorescent Microspheres as *Cryptosporidium parvum* oocyst surrogate
Test Date: February 16th, 2016

Challenge Species	Filter influent average concentration	Average percent removal** of the challenge species by:		
		CUB5k_5 BCS 1602122	CUB5k_6 BCS 1602123	CUB5k_7 BCS 1602124
Bacteria: <i>Raoultella terrigena</i>	6.2 x 10 ⁵ cfu/mL	> 99.9999%*	> 99.9999%*	> 99.9999%*
Virus: MS-2 Bacteriophage	5.4 x 10 ⁵ pfu/mL	99.999%	99.9992%	99.999%
3.0 µM Fluorescent microspheres	4.1 x 10 ⁴ particle/mL	> 99.998%*	> 99.998%*	> 99.998%*

* No species were detected in the filter effluent for the total volume analyzed. Filter effluent samples were analyzed in duplicates at the minimum following collection.

** Purifier NSF/ANSI standard microbial removal claims are 99.9999% or greater for bacteria, 99.99% or greater for virus, and 99.9% or greater for parasite cysts.

BCS LABORATORIES, INC. - GAINESVILLE
 4609 NW 6TH STREET, STE. A, GAINESVILLE, FLORIDA 32609
 TEL. (352) 377-9272, FAX. (352) 377-5630

WWW.MICROBIOSERVICES.COM

FL DOH #E82924, ISO/IEC 17025:2005 L2422 (L-A-B), EPA# FLO1147

FILE: ICON LIFESAVER CUBE FILTER TESTING BCS 1602122-124 02.17.2016

THIS REPORT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN CONSENT OF BCS LABORATORIES



Project: Icon LIFESAVER® cube Efficacy Test
Sample(s): BCS 1602122, 1602123, and 1602124 received February 12th, 2016
Test: Filtration Efficacy – Challenge Test Water (CTW) Type 3
Test Parameter: *Raoultella terrigena* (Bacteria), MS-2 Bacteriophage (virus), and 3.0 µM Fluorescent Microspheres as *Cryptosporidium parvum* oocyst surrogate
Test Date: February 16th, 2016

Challenge Species	Filter influent average concentration	Average concentration of the challenge species in the filters' effluent		
		CUB5k_5 BCS 1602122	CUB5k_6 BCS 1602123	CUB5k_7 BCS 1602124
Bacteria: <i>Raoultella terrigena</i> ¹	5.8 x 10 ⁵ cfu/mL	< 0.45 cfu/mL*	< 0.45 cfu/mL*	< 0.45 cfu/mL*
Virus: MS-2 Bacteriophage ²	4.8 x 10 ⁵ pfu/mL	62.0 pfu/mL	20.9 pfu/mL	54.3 pfu/mL
3.0 µM Fluorescent microspheres ³	4.0 x 10 ⁴ particle/mL	< 1.0 particle/mL*	< 1.0 particle/mL*	< 1.0 particle/mL*

¹ *Raoultella terrigena* (ATCC 33257) was obtained from ATCC and propagated on Tryptic Soy Agar (TSA, Becton Dickinson, USA). It is used to evaluate filters' bacterial removal efficacy. Bacteria was enumerated as colony forming units (cfu) following incubation at 36.5°C for 24 hours as per Standard method 9215C (APHA, 2012).

² Bacteriophage MS-2 (ATCC 15597-B1) was used as a model for human viruses. It is of similar shape and size to human enteroviruses and thus is used to determine filter's viral capture efficacy. It was enumerated using *E. coli* C3000 (ATCC 15597) as a host using the single layer plaque assay agar procedure as per EPA 1601.

³ Three micron green fluorescent latex microspheres (Fluoro-Max™ Green Fluorescent Microspheres 3.00µm, Thermo Scientific CA, USA) were used as surrogates for *Cryptosporidium* oocysts. It is used to determine filter's parasitic removal efficacy. The microspheres were enumerated by fixing onto 3-Well PTFE Slides (Electron Microscopy Sciences, USA) and viewing by UV fluorescence microscopy.

* No species were detected in the filter effluent for the total volume analyzed. Filter effluent samples were analyzed in duplicates at the minimum following collection.

BCS LABORATORIES, INC. - GAINESVILLE
 4609 NW 6TH STREET, STE. A, GAINESVILLE, FLORIDA 32609
 TEL. (352) 377-9272, FAX. (352) 377-5630

WWW.MICROBIOSERVICES.COM

FL DOH #E82924, ISO/IEC 17025:2005 L2422 (L-A-B), EPA# FLO1147

FILE: ICON LIFESAVER CUBE FILTER TESTING BCS 1602122-124 02.17.2016

THIS REPORT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN CONSENT OF BCS LABORATORIES



Project: Icon LIFESAVER® cube Efficacy Test
Sample(s): BCS 1602122, 1602123, and 1602124 received February 12th, 2016
Test: Filtration Efficacy – Challenge Test Water (CTW) Type 3
Test Parameter: Test Parameter: *Raoultella terrigena* (Bacteria), MS-2 Bacteriophage (virus), and 3.0 µM Fluorescent Microspheres as *Cryptosporidium parvum* oocyst surrogate
Test Date: February 16th, 2016

Challenge Species	Filter influent average concentration	Average percent removal** of the challenge species by:		
		CUB5k_5 BCS 1602122	CUB5k_6 BCS 1602123	CUB5k_7 BCS 1602124
Bacteria: <i>Raoultella terrigena</i>	5.8 x 10 ⁵ cfu/mL	> 99.9999%*	> 99.9999%*	> 99.9999%*
Virus: MS-2 Bacteriophage	4.8 x 10 ⁵ pfu/mL	99.99%	99.995%	99.99%
3.0 µM Fluorescent microspheres	4.0 x 10 ⁴ particle/mL	> 99.998%*	> 99.998%*	> 99.998%*

* No species were detected in the filter effluent for the total volume analyzed. Filter effluent samples were analyzed in duplicates at the minimum following collection.

** Purifier NSF/ANSI standard microbial removal claims are 99.9999% or greater for bacteria, 99.99% or greater for virus, and 99.9% or greater for parasite cysts.

BCS LABORATORIES, INC. - GAINESVILLE
 4609 NW 6TH STREET, STE. A, GAINESVILLE, FLORIDA 32609
 TEL. (352) 377-9272, FAX. (352) 377-5630

WWW.MICROBIOSERVICES.COM

FL DOH #E82924, ISO/IEC 17025:2005 L2422 (L-A-B), EPA# FLO1147

FILE: ICON LIFESAVER CUBE FILTER TESTING BCS 1602122-124 02.17.2016

THIS REPORT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN CONSENT OF BCS LABORATORIES



I hereby certify to the accuracy, quality, and data integrity of the reported study. I also certify that the study was appropriately executed and is fully defensible. All physical measurements and their source have been documented. Measurements were obtained using approved protocols and NIST traceable and/or validated instruments. Analysis execution and results were fully documented. Analytical methods used to produce the study's raw data are within the laboratory's ISO 17025 accreditation. The results and conclusions of the study accurately reflect the real raw data obtained in the study.

Signature of Sr. Analyst



David Sekora, M.S.

Date: 02/25/2016



George Lukasik, Ph.D.

Date: 02/25/2016

I certify that I have personally examined and am familiar with the information submitted herein. Based on my inquiry of the individuals immediately responsible for obtaining the information, I certify the submitted information to be true, accurate, and complete. The data provided is solely representative of the analysis conducted on the material/samples/articles provided by the client (or client's representative) it's (their) condition at the time of study. They may not be representative of a process or product. The sample(s) were analyzed in accordance with the method described for each analyte. Due to the inherent limitation(s) of analytical method(s), BCS Laboratories offers no express or implied warranties concerning the quality, safety, and/or purity of any sample, batch, source, or the process they are derived from. The species analysis and corresponding presented results in this report meet the requirements of The NELAC Institute (TNI), ISO 17025, and The State of Florida Department of Public Health's Laboratory Certification Program, as applicable unless otherwise noted.

Signature of Study Director



George Lukasik, Ph.D.

Date: 02/25/2016

- PAGE 13 OF 13 -

BCS LABORATORIES, INC. - GAINESVILLE
4609 NW 6TH STREET, STE. A, GAINESVILLE, FLORIDA 32609
TEL. (352) 377-9272, FAX. (352) 377-5630

WWW.MICROBIOSERVICES.COM

FL DOH #E82924, ISO/IEC 17025:2005 L2422 (L-A-B), EPA# FLO1147

FILE: ICON LIFESAVER CUBE FILTER TESTING BCS 1602122-124 02.17.2016

THIS REPORT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN CONSENT OF BCS LABORATORIES

